



**NATIONAL INSTITUTE FOR RESEARCH AND DEVELOPMENT
IN ELECTRICAL ENGINEERING
INC DIE ICPE-CA**

313 Splaiul Unirii, 030138 Bucharest - 3, ROMANIA

Phone: 0040-21-346.72.31, 346.72.35, 346.82.97, 346.82.98 Fax: 0040-21-346.82.99

<http://www.icpe-ca.ro> e-mail: office@icpe-ca.ro; info@icpe-ca.ro



FILE FOR REQUIREMENT for:

MACHINE FOR SAMPLE PREPARATION

DATE OF SELECTION: February 15, 2008

GENERAL MANAGER,

Prof.dr.Wilhelm KAPPEL

Technical Requirements

Application field: sample preparation for Laser ablation ICP-MS and AAS

Minimal Technical Requirements:

- must be able to obtain samples as solid discs and solutions
- the machine should be automatic or semiautomatic
- the used method : fusion
- must be able to obtain at least 3 samples /hour
- dinamic temperature display
- solid disc dimension: 25-50mm
- the dimensions of the machine must be smaller than width x depth : 100 cm x 70 cm

Financial requirements

- The payment will be from the budget of a National Project, as follow:
 - 30% of the value in 10 days after the signing of the contract
 - 70 % of the value in 30 days after the reception

Eligibility requirement:

- Declaration of compliance;
- Certificate of the payment of all taxes to the national authorities;
- Certificate of incorporation the company in their homeland;
- Certificate of allotment for the equipment;
- General information;
- Recommendation from other beneficiaries/customers;
- The last Balance sheet endorsed and registered by competent institution.
- The offer is valid 60 days from the date of fall out.

The criteria used for choose: the best technical and economical offer

Criteria	Points
1. technical characteristics	50
2. price	30
3. delivery terms	10
4. payment	10
TOTAL	100